


Inside the July Issue



Click the cover above for the online edition, or click below for the digital flip book.



July Cover Story

[Physician preference: Supply Chain matters for patient, industry health](#)

HPN 2016 INDUSTRY GUIDE

[Self Study Series](#)

[White Papers](#)

[Special Reports](#)

[Purchasing Connection](#)

[Resources](#)

[Show Calendar](#)

[HPN Hall of Fame](#)

[HPN Buyers Guides](#)

SHOP HPN
FEATURED PRODUCTS
INFORMATION LINK

[Classifieds](#)

[Issue Archives](#)

[Advertise](#)

[About Us](#)

MON TASTIC

HEALTHCARE PURCHASING NEWS

Follow @HPN_Online

CLINICAL INTELLIGENCE FOR SUPPLY CHAIN LEADERSHIP

Like us on Facebook

INSIDE THE CURRENT ISSUE

July 2016

CS Solutions



Submit your questions:

email: editor@hpnonline.com

phone: (941) 927-9345 ext.202.

Medical grade air; dry vs. moist instrument transport; multiple inspection magnifiers

by Ray Taurasi

Q I am an infection control manager and recently attended an infection control conference. One of the speakers talked about various utility requirements for the sterile processing area and had emphasized the need to have medical grade air readily available. I don't quite understand exactly what medical grade air is or its application in sterile processing. Does this apply to the environmental air that is circulated through the department or is it oxygen for emergency carts?

A Medical grade air, sometimes referred to as instrument air, is basically a medical gas that must meet the requirements set forth by the National Fire Protection Association, NFPA 99: Health Care Facilities Code. This air is not respired and must be compliant with American National Standards Institute and International Society of Automation ANSI/ISA S-7.0.01. That is the quality standard for instrument air, which is filtered to 0.01 micron, free of liquids, and hydrocarbon vapors and dry to a dew point of -40°F (4.444°C). There are medical devices such as those with lumens that require drying and instrument air is blown through the channels; instrument air may also be used for testing the performance of some medical devices. (See Figure 1.)

Q I am the perioperative educator and quality assurance specialist for our offsite ambulatory care and surgery centers. During a recent accreditation inspection the surveyor issued a citation because cleaned instruments were being returned to the main hospital dry and in a closed tote container. The surveyor referenced an infection control standard that said instrumentation should be kept moist during transportation for final processing. I questioned the citation and told her our

Visit our
CS Solutions Archive

Trade your syringes in for a **faster** and **easier** method of pre-cleaning scopes!

pure PROCESSING

[LEARN MORE ▶](#)



Data should fuel our humanity, not distract from it.

vizient.

[Click to connect with us.](#)

AHRMM16
COLLABORATING TO ACHIEVE THE TRIPLE AIM
WHERE HEALTHCARE SUPPLY CHAIN MEETS
July 31 - August 3
San Diego
AHRMM16.ORG

Current Issue JULY 2016

[Contact Us](#)[Subscribe](#)

Receive our



HEALTHCARE
PURCHASING NEWS
Daily Update
e-Newsletter

[Sign Up Now](#)Privacy by  SafeSubscribeSM[HOME](#)

KSR Publishing, Inc.
Copyright © 2016

policy was for the satellite centers to manually clean all instruments, in accordance with IFUs, and then place them in a covered bin for return to the main hospital where they are packaged and sterilized. The surveyor held firm on the citation and further recommended that we assess our entire process for the handling and transportation of instrumentation between the satellite locations and the main hospital campus. I really feel that the citation is unwarranted and believe our method of handling "clean" instrumentation is appropriate. Am I wrong? What am I missing here?

A Without knowing the standard, source and reference for the citation I cannot really fully address or respond to your issues and concerns. In general, I can see no rationale for or need to keep clean instruments moist during transportation. There are several recommendations, guidelines and standards relative to the transportation and handling of soiled, clean and sterile medical devices. The accrediting bodies and their surveyors are very watchful and concerned regarding proper handling and transportation of medical devices in accordance with standards of practice. It seems to me that the surveyor may not have fully understood your protocol or likely saw practices that were questionable or conflicting. You should be able to obtain the reference standard for which you were cited as non-compliant in order to respond with a corrective action plan.

When assessing your current protocol you should include the following:

- Does your process allow for complete segregation of clean, sterile and soiled items from one another?
- Are all items properly contained in covered, rigid, puncture-proof containers?
- Are all containers clearly labeled and identifiable as clean, sterile, soiled and/or biohazardous?
- Are the conditions of the soiled and clean processing rooms in compliance with standards?
- Are the manual cleaning procedures strictly enforced and in compliance with the device manufacturer's IFUs?
- Are all the personnel responsible for processing /cleaning the instruments at the various satellite centers trained and competent to perform these functions?
- Are personnel following the proper use of personal protective equipment (PPE)?
- Are the processing protocols consistent throughout the organization?
- Are the transportation vehicles in compliance with various DPH, DOT regulations?
- Are policies and procedures readily accessible to personnel at all locations?
- Are quality assurance protocols for cleaning verification and documentation in place?

Q I am the lead Central Sterile Processing (CSP) technician and report to the OR nurse manager. We only have one magnifier in our department which I know is not sufficient. I would like to have one at every work station and we really need a borescope to inspect our shaves and small instruments

[3M](#)[Acute Care Pharmaceuticals](#)[AHRMM](#)[Ansell Professional Healthcare](#)[BD Genesis](#)[Belimed](#)[Boston Scientific](#)[Cardinal Health - ASC Solutions](#)[Cardinal Health - OptiFreight](#)[Cardinal Health - PPE](#)[Construction Specialties](#)[Contec Inc.](#)[CS Medical](#)[Dale Medical Products Inc.](#)[Detecto](#)[Healthmark Industries](#)[HealthTrust Purchasing Group](#)[IMS](#)[Key Surgical](#)[Mölnlycke - Biogel](#)[Mölnlycke - Hibiclens](#)[PDI](#)[Pegasus Medical](#)[Pure Processing](#)[Ruhof Corp. - Scope Products](#)[Ruhof Corp. - CleanStart](#)[Ryder System Inc.](#)[Sage Products](#)[Sherwin-Williams](#)[TRIOSE](#)[Tronex Healthcare Industries](#)[Tru-D](#)[Vizient](#)

with channels. Every time I bring this up to my manager she says the one we have should be enough and that there is no requirement to have one at every station. She says that the budget is tight and that there is no way we can spend several thousand dollars on magnifiers. What can I do to convince her that we do need more inspection devices?

A You are absolutely correct - one magnifier is not enough to conduct the inspections of your instruments and other medical devices properly. There are many different types of visual inspection tools that should be used in CS to thoroughly inspect devices. (See Figure 2.) One type of visual inspection device cannot meet all of your needs. Most instrument manufacturers' IFUs include a visual inspection step. Inspection is of course necessary to ensure that visual soils are removed but instruments also need to be inspected to be certain they are not damaged. Many instruments have small parts, recessed crevices, channels and other areas that cannot be seen by the naked eye or reached without the use of a borescope.

Just about all of the professional organizations that have guidelines or standards including AAMI, AORN, SGNA, emphasize the need for visual inspection. I would suggest that you review these documents for support as well as the IFUs of your instrument manufacturers, especially orthopedic devices such as arthroscopic shavers. To help get you get started here is an excerpt from ANSI / AAMI ST79 Section -7.5.5: "After completing the cleaning process, personnel should visually inspect each item carefully to detect any visible soil. Inspection using magnification might identify residues more readily than the unaided eye."

I recommend that you gather all of the information necessary to support your need for additional visual inspection tools and share it with your manager, along with a request for the devices you need. [HPN](#)

Ray Taurasi is Eastern Regional Director of Clinical Sales and Services for [Healthmark Industries](#). His healthcare career spans over three decades as an Administrator, Educator, Technologist and Consultant. He is a member of AORN, AHA, SGNA, AAMI and a past president of IAHCSSM and has served on and contributed to many national committees with a myriad of professional organizations, manufacturers, corporations and prestigious healthcare networks. Taurasi has been a faculty member of numerous colleges teaching in the divisions of business administration and health sciences. In addition to this column he has authored several articles and has been a featured speaker on the international scene.

